



**The Marine Products Export Development Authority  
(Ministry of Commerce & Industry, Govt. of India)**

Head Office, MPEDA House, Building No: 27/1162, PB No:4272,  
Panampilly Avenue, Panampilly Nagar PO, KOCHI-682036, KERALA  
Telephone(EPABX):+91484 2311979 Fax:+914842313361/  
E-mail:[ho@mpeda.gov.in](mailto:ho@mpeda.gov.in), [lab.koc@mpeda.gov.in](mailto:lab.koc@mpeda.gov.in)  
Web:[www.mpeda.gov.in](http://www.mpeda.gov.in)

फासं/File No: LAB-KOC/PUR(GAS)/1/2022-LAB

दिनांक/Date: 06/01/2023

**CORRIGENDUM NO: 1**

**TENDER ID: 2022\_MPEDA\_730960\_1**

With reference to the tender published by MPEDA for the refilling and supply of various gas cylinders to MPEDA Quality Control Laboratory at Kochi vide Tender Ref. No. **LAB-KOC/PUR(GAS)/1/2022-LAB** dated 18.11.2022 and TENDER ID: 2022\_MPEDA\_730960\_1, following to be noted:

Few amendments are made to technical specifications on gas purity. The revised specifications are given below:

Sl. No.	Existing Item and Specification/purity	Revised specification
1	UHP Helium 99.9999%, O2:<2.0 ppm N2:<4.0 ppm CO+CO2:<0.5 ppm THC:<0.5 ppm H2O:<2.0 ppm	UHP Helium <b>99.999%</b> , O2:< <b>1.0</b> ppm N2:<4.0 ppm CO+CO2:<0.5 ppm THC:<0.5 ppm H2O:< <b>1.0</b> ppm
2	Argon Gas 99.996% O2:<2.0 ppm CO+CO2:<0.5 ppm THC:<0.5 ppm H2O:<5.0 ppm	Argon Gas 99.996% <b>N2:&lt;4.0 ppm</b> O2:<2.0 ppm CO+CO2:<0.5 ppm THC:<0.5 ppm H2O:<5.0 ppm

All other terms and conditions in the tender remain unchanged.

**Joint Director (QC)**  
**The Marine Products Export Development Authority**  
**MPEDA House, Panampilly Avenue,**  
**Kochi-682036, Kerala**